



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/543,083	01/04/2006	Mark Humayun	58478 (50420)	2575
21874 7590 11/25/2009 EDWARDS ANGELL PALMER & DODGE LLP P.O. BOX 55874 BOSTON, MA 02205				
EXAMINER				
MENDEZ, MANUEL A				
ART UNIT		PAPER NUMBER		
3763				
MAIL DATE		DELIVERY MODE		
11/25/2009		PAPER		

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

### Office Action Summary

**Application No.**

10/543,083

**Applicant(s)**

HUMAYUN ET AL.

**Examiner**

Manuel A. Mendez

**Art Unit**

3763

**Period for Reply** -- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**Status**

- 1) ☒ Responsive to communication(s) filed on 11 August 2009.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

**Disposition of Claims**

- 4) ☒ Claim(s) 1-6 and 8-23 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☒ Claim(s) 19-22 is/are allowed.
- 6) ☒ Claim(s) 1-6, 8-18 and 23 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

**Application Papers**

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

**Priority under 35 U.S.C. § 119**

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

**Attachment(s)**

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO/S508)  
Paper No(s)/Mail Date \_\_\_\_\_
- 4) ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date \_\_\_\_\_
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: \_\_\_\_\_

## **DETAILED ACTION**

### ***Examiner's Comment***

The examiner of record noticed that the last office action was intended to be a final action; however, the forms were not properly marked to indicate the finality of the action and the closing of the prosecution. Accordingly, in order to expedite the prosecution of this application, this supplemental action is presented to fix the deficiencies of the last office action. Please note that in view of the noted discrepancies, this action is not a final action.

### ***Response to Amendment***

In response to applicant's argument indicating that "Avery does not teach or suggest a subretinal delivery device having a cannula configured to extend from a reservoir through the retina into a subretinal space when the reservoir is disposed exterior the eye on the sclera or when the reservoir is disposed within the eye, such that an agent in the reservoir is released through the cannula to the eye subretinally", the examiner notes for the record that a recitation of the intended use of the claimed invention must result in a structural difference between the claimed invention and the prior art in order to patentably distinguish the claimed invention from the prior art. If the prior art structure is capable of performing the intended use, then it meets the claim. In this particular case, the cited patents disclose the structural elements disclosed in the claims in question, and therefore, it is concluded that the rejections to claims 1-18 and 23 are merited and will be maintained.

### ***Claim Rejections - 35 USC § 102***

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

**Claim 1** is rejected under 35 U.S.C. 102(b) as being anticipated by **Avery et al. (5725493)**. The cited patent discloses a reservoir, a cannula extending from the reservoir and wherein an agent stored in the reservoir is released through the cannula.

***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

**Claims 2-18 and 23** are rejected under 35 U.S.C. 103(a) as being unpatentable over **Avery et al.** in view of **Krupin (5,454,796)** or **Memmen (5,370,607)**. The **Avery et al.** patent does not disclose a ring-like delivery device. However, such shape designs are conventional in the art as evidenced by the teachings of **Krupin** or **Memmen**.

**Krupin** shows in figure 5 a ring-shaped delivery device. Additionally, **Memmen** also shows in figure 3, a delivery device shaped as a ring. Based on the above observations, for a person of ordinary skill in the art, it would have been obvious to modify the shape of the delivery device disclosed by **Avery et al.** with a ring shape as taught by the **Krupin** or **Memmen** patents in view of the proven conventionality of the ring-shape modification. Moreover, for an artisan skilled in the art, such shape modification would have been considered another alternative in the process of designing a delivery device apparatus.

***Allowable Subject Matter***

**Claims 19-22** are objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

***Conclusion***

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Manuel A. Mendez whose telephone number is 571-272-4962. The examiner can normally be reached on 0730-1800 hrs.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Mr. Nicholas D. Lucchesi can be reached on 571-272-4977. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Manuel A. Mendez/

Primary Examiner, Art Unit 3763

Manuel A. Mendez  
Primary Examiner  
Art Unit 3763

MM